

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG090721\
 Data File : FG008732.D
 Signal(s) : FID1A.ch
 Acq On : 07 Sep 2021 15:28
 Operator : DD\AJ
 Sample : 10 TRPH STD
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Sep 08 02:35:33 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG090721.M
 Quant Title :
 QLast Update : Wed Sep 08 02:33:45 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.745	1126079	9.844 ug/ml
Target Compounds			
1) N-OCTANE	2.346	1311205	9.530 ug/ml
2) N-DECANE	4.447	1327543	9.475 ug/ml
3) N-DODECANE	6.473	1345397	9.478 ug/ml
4) N-TETRADECANE	8.275	1349681	9.472 ug/ml
5) N-HEXADECANE	9.881	1330286	9.488 ug/ml
6) N-OCTADECANE	11.326	1345513	9.616 ug/ml
7) N-EICOSANE	12.640	1338435	9.754 ug/ml
8) N-DOCOSANE	13.842	1301798	9.810 ug/ml
10) N-TETRACOSANE	14.950	1280277	9.826 ug/ml
11) N-HEXACOSANE	15.973	1245543	9.865 ug/ml
12) N-OCTACOSANE	16.926	1223940	10.052 ug/ml
13) N-TRIACONTANE	17.815	1175024	10.196 ug/ml
14) N-DOTRIACONTANE	18.649	1085823	10.236 ug/ml
15) N-TETRATRIACONTANE	19.435	982871	9.978 ug/ml
16) N-HEXATRIACONTANE	20.177	912462	9.816 ug/ml
17) N-OCTATRIACONTANE	20.931	881560	9.799 ug/ml
18) N-TETRACONTANE	21.877	915948	9.918 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG090721\
Data File : FG008732.D
Signal(s) : FID1A.ch
Acq On : 07 Sep 2021 15:28
Operator : DD\AJ
Sample : 10 TRPH STD
Misc :
ALS Vial : 6 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Sep 08 02:35:33 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG090721.M
Quant Title :
QLast Update : Wed Sep 08 02:33:45 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

