

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG090722\
 Data File : FG010520.D
 Signal(s) : FID1A.ch
 Acq On : 07 Sep 2022 17:28
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: Sep 08 02:47:22 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG082622.M
 Quant Title :
 QLast Update : Fri Aug 26 23:48:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.783	5589470	45.671 ug/ml
Target Compounds			
2) N-DECANE	4.312	6354289	49.574 ug/ml
3) N-DODECANE	6.493	6472068	49.176 ug/ml
4) N-TETRADECANE	8.322	6504613	48.058 ug/ml
5) N-HEXADECANE	9.929	6615911	47.215 ug/ml
6) N-OCTADECANE	11.374	6658301	46.635 ug/ml
7) N-EICOSANE	12.683	6603999	46.253 ug/ml
8) N-DOCOSANE	13.882	6410447	46.002 ug/ml
10) N-TETRACOSANE	14.986	6336599	45.731 ug/ml
11) N-HEXACOSANE	16.008	6166709	45.457 ug/ml
12) N-OCTACOSANE	16.958	5983627	45.247 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG090722\
Data File : FG010520.D
Signal(s) : FID1A.ch
Acq On : 07 Sep 2022 17:28
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
50 PPM TRPH STD

Integration File: autoint1.e
Quant Time: Sep 08 02:47:22 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG082622.M
Quant Title :
QLast Update : Fri Aug 26 23:48:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

