

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG090822\
 Data File : FG010533.D
 Signal(s) : FID1A.ch
 Acq On : 08 Sep 2022 19:39
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 50 PPM TRPH STD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/09/2022
 Supervised By :Ankita Jodhani 09/09/2022

Integration File: Sample.e
 Quant Time: Sep 09 07:56:32 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG082622.M
 Quant Title :
 QLast Update : Fri Aug 26 23:48:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.787	5660699	46.253 ug/ml
Target Compounds			
1) N-OCTANE	1.809	6272086	49.839 ug/ml
2) N-DECANE	4.313	6586234	51.383 ug/ml
3) N-DODECANE	6.494	6648499	50.517 ug/ml
4) N-TETRADECANE	8.325	6690654	49.432 ug/ml
5) N-HEXADECANE	9.933	6768317	48.303 ug/ml
6) N-OCTADECANE	11.376	6799176	47.621 ug/ml
7) N-EICOSANE	12.687	6715535	47.034 ug/ml
8) N-DOCOSANE	13.886	6499461	46.640 ug/ml
10) N-TETRACOSANE	14.990	6416835	46.310 ug/ml
11) N-HEXACOSANE	16.012	6220518	45.853 ug/ml
12) N-OCTACOSANE	16.962	6053285	45.774 ug/ml
13) N-TRIACONTANE	17.851	5967686	46.055 ug/ml
14) N-DOTRIACONTANE	18.686	5937719	46.443 ug/mlm
15) N-TETRATRIACONTANE	19.469	5956660	48.151 ug/mlm
16) N-HEXATRIACONTANE	20.209	5768126	49.696 ug/mlm
17) N-OCTATRIACONTANE	20.923	6041915	52.604 ug/mlm
18) N-TETRACONTANE	21.763	6779506	57.502 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG090822\
Data File : FG010533.D
Signal(s) : FID1A.ch
Acq On : 08 Sep 2022 19:39
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
50 PPM TRPH STD

Manual Integrations
APPROVED

Reviewed By : Yogesh Patel 09/09/2022
Supervised By : Ankita Jodhani 09/09/2022

Integration File: Sample.e
Quant Time: Sep 09 07:56:32 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG082622.M
Quant Title :
QLast Update : Fri Aug 26 23:48:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

