

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG091319\
Data File : FG005938.D
Signal(s) : FID1A.ch
Acq On : 14 Sep 2019 00:01
Operator : DD\AJ
Sample : K4663-04 5X
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
OK-01-091119

Integration File: autoint1.e
Quant Time: Sep 14 04:12:36 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG090719.M
Quant Title :
QLast Update : Mon Sep 09 05:03:58 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.726	273307	2.433 ug/ml
Target Compounds			
10) N-TETRACOSANE	14.929	677611	5.313 ug/ml
12) N-OCTACOSANE	16.942	192925	1.636 ug/ml
14) N-DOTRIACONTANE	18.618	419989	3.924 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG091319\
Data File : FG005938.D
Signal(s) : FID1A.ch
Acq On : 14 Sep 2019 00:01
Operator : DD\AJ
Sample : K4663-04 5X
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
OK-01-091119

Integration File: autoint1.e
Quant Time: Sep 14 04:12:36 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG090719.M
Quant Title :
QLast Update : Mon Sep 09 05:03:58 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um






#9 TETRACOSANE-d50 (SURROGATE)


R.T.: 14.726 min
Delta R.T.: -0.005 min
Response: 273307
Conc: 2.43 ug/ml

Instrument :
FID_G
ClientSampleId :
OK-01-091119




#10 N-TETRACOSANE

R.T.: 14.929 min
Delta R.T.: -0.006 min
Response: 677611
Conc: 5.31 ug/ml



#12 N-OCTACOSANE

R.T.: 16.942 min
Delta R.T.: 0.033 min
Response: 192925
Conc: 1.64 ug/ml



#14 N-DOTRIACONTANE

R.T.: 18.618 min
Delta R.T.: -0.014 min
Response: 419989
Conc: 3.92 ug/ml