

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG091918\
 Data File : FG002337.D
 Signal(s) : FID1A.ch
 Acq On : 19 Sep 2018 10:49
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Sep 20 03:57:58 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG083018.M
 Quant Title :
 QLast Update : Thu Aug 30 14:49:31 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.425	6293703	49.886 ug/ml
Target Compounds			
2) N-DECANE	4.174	7322302	61.002 ug/ml
3) N-DODECANE	6.188	7477195	59.415 ug/ml
4) N-TETRADECANE	7.980	7538788	57.425 ug/ml
5) N-HEXADECANE	9.579	7567823	55.600 ug/ml
6) N-OCTADECANE	11.019	7530745	53.613 ug/ml
7) N-EICOSANE	12.328	7394661	52.050 ug/ml
8) N-DOCOSANE	13.526	7242527	50.747 ug/ml
10) N-TETRACOSANE	14.628	7079047	49.517 ug/ml
11) N-HEXACOSANE	15.650	6948005	48.393 ug/ml
12) N-OCTACOSANE	16.599	6865000	46.920 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG091918\
Data File : FG002337.D
Signal(s) : FID1A.ch
Acq On : 19 Sep 2018 10:49
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Sep 20 03:57:58 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG083018.M
Quant Title :
QLast Update : Thu Aug 30 14:49:31 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

