

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG091922\
 Data File : FG010602.D
 Signal(s) : FID1A.ch
 Acq On : 19 Sep 2022 10:27
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Sep 20 03:19:48 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG082622.M
 Quant Title :
 QLast Update : Fri Aug 26 23:48:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
9) S	TETRACOSANE-d50 (SURR...	14.785	6103317	49.870 ug/ml
Target Compounds				
2)	N-DECANE	4.306	7145374	55.746 ug/ml
3)	N-DODECANE	6.490	7279760	55.313 ug/ml
4)	N-TETRADECANE	8.322	7347192	54.283 ug/ml
5)	N-HEXADECANE	9.931	7471398	53.320 ug/ml
6)	N-OCTADECANE	11.374	7491932	52.473 ug/ml
7)	N-EICOSANE	12.686	7380639	51.692 ug/ml
8)	N-DOCOSANE	13.885	7082800	50.826 ug/ml
10)	N-TETRACOSANE	14.989	6903080	49.820 ug/ml
11)	N-HEXACOSANE	16.012	6651233	49.028 ug/ml
12)	N-OCTACOSANE	16.961	6572619	49.701 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG091922\
Data File : FG010602.D
Signal(s) : FID1A.ch
Acq On : 19 Sep 2022 10:27
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Sep 20 03:19:48 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG082622.M
Quant Title :
QLast Update : Fri Aug 26 23:48:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

