

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG091922\
 Data File : FG010613.D
 Signal(s) : FID1A.ch
 Acq On : 19 Sep 2022 16:03
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Sep 20 03:21:09 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG082622.M
 Quant Title :
 QLast Update : Fri Aug 26 23:48:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.783	6356079	51.935 ug/ml
Target Compounds			
2) N-DECANE	4.309	7534409	58.781 ug/ml
3) N-DODECANE	6.490	7685890	58.399 ug/ml
4) N-TETRADECANE	8.321	7738904	57.177 ug/ml
5) N-HEXADECANE	9.928	7843972	55.979 ug/ml
6) N-OCTADECANE	11.373	7832475	54.858 ug/ml
7) N-EICOSANE	12.683	7692116	53.874 ug/ml
8) N-DOCOSANE	13.882	7378793	52.951 ug/ml
10) N-TETRACOSANE	14.986	7196340	51.936 ug/ml
11) N-HEXACOSANE	16.009	6934494	51.116 ug/ml
12) N-OCTACOSANE	16.961	6808024	51.481 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG091922\
Data File : FG010613.D
Signal(s) : FID1A.ch
Acq On : 19 Sep 2022 16:03
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Sep 20 03:21:09 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG082622.M
Quant Title :
QLast Update : Fri Aug 26 23:48:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

