

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG092319\
 Data File : FG006021.D
 Signal(s) : FID1A.ch
 Acq On : 23 Sep 2019 12:17
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Sep 24 00:50:28 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG090719.M
 Quant Title :
 QLast Update : Mon Sep 09 05:03:58 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.735	5253834	46.764 ug/ml
Target Compounds			
2) N-DECANE	4.431	7510482	53.513 ug/ml
3) N-DODECANE	6.460	7341061	51.848 ug/ml
4) N-TETRADECANE	8.263	7068670	50.135 ug/ml
5) N-HEXADECANE	9.870	6865555	48.851 ug/ml
6) N-OCTADECANE	11.316	6558923	47.825 ug/ml
7) N-EICOSANE	12.631	6431696	47.370 ug/ml
8) N-DOCOSANE	13.834	6268289	47.654 ug/ml
10) N-TETRACOSANE	14.941	6258374	49.074 ug/ml
11) N-HEXACOSANE	15.965	6249675	51.000 ug/ml
12) N-OCTACOSANE	16.918	6335139	53.722 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG092319\
Data File : FG006021.D
Signal(s) : FID1A.ch
Acq On : 23 Sep 2019 12:17
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Sep 24 00:50:28 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG090719.M
Quant Title :
QLast Update : Mon Sep 09 05:03:58 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

