

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG092519\
 Data File : FG006049.D
 Signal(s) : FID1A.ch
 Acq On : 25 Sep 2019 19:41
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Sep 26 01:25:43 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG090719.M
 Quant Title :
 QLast Update : Mon Sep 09 05:03:58 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.733	5493945	48.901 ug/ml
Target Compounds			
2) N-DECANE	4.429	7254296	51.687 ug/ml
3) N-DODECANE	6.458	7126704	50.334 ug/ml
4) N-TETRADECANE	8.261	6896226	48.912 ug/ml
5) N-HEXADECANE	9.868	6736440	47.933 ug/ml
6) N-OCTADECANE	11.314	6478295	47.237 ug/ml
7) N-EICOSANE	12.629	6417737	47.267 ug/ml
8) N-DOCOSANE	13.832	6324826	48.083 ug/ml
10) N-TETRACOSANE	14.938	6406471	50.235 ug/ml
11) N-HEXACOSANE	15.963	6386991	52.120 ug/ml
12) N-OCTACOSANE	16.918	6484239	54.986 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG092519\
Data File : FG006049.D
Signal(s) : FID1A.ch
Acq On : 25 Sep 2019 19:41
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Sep 26 01:25:43 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG090719.M
Quant Title :
QLast Update : Mon Sep 09 05:03:58 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

