

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG092922\
 Data File : FG010725.D
 Signal(s) : FID1A.ch
 Acq On : 29 Sep 2022 20:10
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: Sep 30 04:21:55 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG092822.M
 Quant Title :
 QLast Update : Thu Sep 29 02:21:49 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.753	5523248	45.526 ug/ml
Target Compounds			
2) N-DECANE	4.273	6371154	48.957 ug/ml
3) N-DODECANE	6.456	6534296	49.271 ug/ml
4) N-TETRADECANE	8.288	6583396	49.096 ug/ml
5) N-HEXADECANE	9.896	6678078	48.162 ug/ml
6) N-OCTADECANE	11.340	6680293	47.399 ug/ml
7) N-EICOSANE	12.652	6592930	46.623 ug/ml
8) N-DOCOSANE	13.851	6365992	45.987 ug/ml
10) N-TETRACOSANE	14.958	6256400	45.411 ug/ml
11) N-HEXACOSANE	15.978	6070436	44.995 ug/ml
12) N-OCTACOSANE	16.933	5887062	44.893 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG092922\
Data File : FG010725.D
Signal(s) : FID1A.ch
Acq On : 29 Sep 2022 20:10
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
50 PPM TRPH STD

Integration File: autoint1.e
Quant Time: Sep 30 04:21:55 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG092822.M
Quant Title :
QLast Update : Thu Sep 29 02:21:49 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

