

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG100218\
 Data File : FG002481.D
 Signal(s) : FID1A.ch
 Acq On : 02 Oct 2018 19:33
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Oct 03 05:31:36 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG100118.M
 Quant Title :
 QLast Update : Mon Oct 01 12:36:17 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.758	5980795	52.106 ug/ml
Target Compounds			
1) N-OCTANE	2.342	6642687	52.171 ug/ml
2) N-DECANE	4.462	6912897	53.057 ug/ml
3) N-DODECANE	6.492	6988891	52.958 ug/ml
4) N-TETRADECANE	8.292	7033481	52.751 ug/ml
5) N-HEXADECANE	9.897	7062224	52.646 ug/ml
6) N-OCTADECANE	11.342	7076906	52.548 ug/ml
7) N-EICOSANE	12.656	6984722	52.363 ug/ml
8) N-DOCOSANE	13.858	6881281	52.149 ug/ml
10) N-TETRACOSANE	14.963	6761947	52.014 ug/ml
11) N-HEXACOSANE	15.985	6630054	51.613 ug/ml
12) N-OCTACOSANE	16.937	6523790	51.456 ug/ml
13) N-TRIACONTANE	17.825	6479433	51.494 ug/ml
14) N-DOTRIACONTANE	18.659	6356091	51.771 ug/ml
15) N-TETRATRIACONTANE	19.446	6274473	52.583 ug/ml
16) N-HEXATRIACONTANE	20.188	6012979	54.197 ug/ml
17) N-OCTATRIACONTANE	20.949	5421229	55.305 ug/ml
18) N-TETRACONTANE	21.901	5342934	57.125 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG100218\
Data File : FG002481.D
Signal(s) : FID1A.ch
Acq On : 02 Oct 2018 19:33
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Oct 03 05:31:36 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG100118.M
Quant Title :
QLast Update : Mon Oct 01 12:36:17 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

