

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG100220\
 Data File : FG007164.D
 Signal(s) : FID1A.ch
 Acq On : 02 Oct 2020 22:35
 Operator : DD\AJ
 Sample : 5 TRPH STD
 Misc :
 ALS Vial : 62 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Oct 05 12:50:41 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG100220.M
 Quant Title :
 QLast Update : Mon Oct 05 12:49:05 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.739	644342	5.533 ug/ml
Target Compounds			
1) N-OCTANE	2.340	669533	5.142 ug/ml
2) N-DECANE	4.445	683793	5.149 ug/ml
3) N-DODECANE	6.471	690758	5.154 ug/ml
4) N-TETRADECANE	8.271	685275	5.133 ug/ml
5) N-HEXADECANE	9.877	717445	5.243 ug/ml
6) N-OCTADECANE	11.323	734052	5.334 ug/ml
7) N-EICOSANE	12.635	740464	5.446 ug/ml
8) N-DOCOSANE	13.837	735465	5.494 ug/ml
10) N-TETRACOSANE	14.943	738821	5.530 ug/ml
11) N-HEXACOSANE	15.967	730283	5.574 ug/ml
12) N-OCTACOSANE	16.918	726100	5.644 ug/ml
13) N-TRIACONTANE	17.807	734132	5.804 ug/ml
14) N-DOTRIACONTANE	18.640	705859	5.917 ug/ml
15) N-TETRATRIACONTANE	19.424	598614	5.613 ug/ml
16) N-HEXATRIACONTANE	20.166	474363	5.260 ug/ml
17) N-OCTATRIACONTANE	20.918	412155	5.083 ug/ml
18) N-TETRACONTANE	21.859	338011	4.244 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG100220\
Data File : FG007164.D
Signal(s) : FID1A.ch
Acq On : 02 Oct 2020 22:35
Operator : DD\AJ
Sample : 5 TRPH STD
Misc :
ALS Vial : 62 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Oct 05 12:50:41 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG100220.M
Quant Title :
QLast Update : Mon Oct 05 12:49:05 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

