

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG100318\
 Data File : FG002501.D
 Signal(s) : FID1A.ch
 Acq On : 03 Oct 2018 22:39
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Oct 04 05:33:32 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG100118.M
 Quant Title :
 QLast Update : Mon Oct 01 12:36:17 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.759	5655290	49.270 ug/ml
Target Compounds			
2) N-DECANE	4.463	6627876	50.869 ug/ml
3) N-DODECANE	6.492	6689196	50.687 ug/ml
4) N-TETRADECANE	8.293	6715310	50.365 ug/ml
5) N-HEXADECANE	9.898	6733793	50.197 ug/ml
6) N-OCTADECANE	11.345	6731263	49.982 ug/ml
7) N-EICOSANE	12.658	6628247	49.691 ug/ml
8) N-DOCOSANE	13.858	6502578	49.279 ug/ml
10) N-TETRACOSANE	14.964	6345087	48.808 ug/ml
11) N-HEXACOSANE	15.986	6217007	48.398 ug/ml
12) N-OCTACOSANE	16.939	6122809	48.293 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG100318\
Data File : FG002501.D
Signal(s) : FID1A.ch
Acq On : 03 Oct 2018 22:39
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Oct 04 05:33:32 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG100118.M
Quant Title :
QLast Update : Mon Oct 01 12:36:17 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

