

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG100318\
Data File : FG002501.D
Signal(s) : FID1A.ch
Acq On : 03 Oct 2018 22:39
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG100118.M
Title :

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	4.463	4.431	4.501	BB	731011	6627876	98.43%	9.339%
2	6.492	6.454	6.524	BB	686582	6689196	99.34%	9.426%
3	8.293	8.254	8.329	BB	660784	6715310	99.73%	9.462%
4	9.898	9.852	9.941	BB	633449	6733793	100.00%	9.488%
5	11.345	11.304	11.387	BB	590948	6731263	99.96%	9.485%
6	12.658	12.617	12.699	BB	556006	6628247	98.43%	9.340%
7	13.858	13.819	13.902	BB	523799	6502578	96.57%	9.163%
8	14.759	14.716	14.832	BB	408795	5655290	83.98%	7.969%
9	14.964	14.924	15.014	BB	521264	6345087	94.23%	8.941%
10	15.986	15.942	16.037	BB	511595	6217007	92.33%	8.760%
11	16.939	16.877	16.991	BB	449083	6122809	90.93%	8.628%
Sum of corrected areas:						70968457		

FG100118.M Thu Oct 04 08:39:32 2018