

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG100319\
 Data File : FG006129.D
 Signal(s) : FID1A.ch
 Acq On : 03 Oct 2019 15:30
 Operator : DD\AJ
 Sample : 20 TRPH STD
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Oct 04 04:30:50 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG100319.M
 Quant Title :
 QLast Update : Fri Oct 04 04:29:25 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.474	1977652	20.069 ug/ml
Target Compounds			
1) N-OCTANE	2.277	2175441	21.207 ug/ml
2) N-DECANE	4.340	2270666	21.171 ug/ml
3) N-DODECANE	6.335	2328239	20.947 ug/ml
4) N-TETRADECANE	8.106	2345631	20.635 ug/ml
5) N-HEXADECANE	9.686	2369559	20.386 ug/ml
6) N-OCTADECANE	11.111	2344665	20.191 ug/ml
7) N-EICOSANE	12.405	2344797	20.101 ug/ml
8) N-DOCOSANE	13.587	2299671	20.101 ug/ml
10) N-TETRACOSANE	14.676	2256137	20.187 ug/ml
11) N-HEXACOSANE	15.683	2222101	20.402 ug/ml
12) N-OCTACOSANE	16.619	2253235	20.928 ug/ml
13) N-TRIACONTANE	17.494	2196770	21.069 ug/ml
14) N-DOTRIACONTANE	18.316	1998541	20.662 ug/ml
15) N-TETRATRIACONTANE	19.089	1910848	20.352 ug/ml
16) N-HEXATRIACONTANE	19.820	1793493	20.419 ug/ml
17) N-OCTATRIACONTANE	20.517	1650150	20.247 ug/ml
18) N-TETRACONTANE	21.318	1671458	20.479 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG100319\
Data File : FG006129.D
Signal(s) : FID1A.ch
Acq On : 03 Oct 2019 15:30
Operator : DD\AJ
Sample : 20 TRPH STD
Misc :
ALS Vial : 5 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Oct 04 04:30:50 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG100319.M
Quant Title :
QLast Update : Fri Oct 04 04:29:25 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

