

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG100319\
 Data File : FG006132.D
 Signal(s) : FID1A.ch
 Acq On : 03 Oct 2019 16:59
 Operator : DD\AJ
 Sample : FG100319ICV
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Oct 04 04:37:00 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG100319.M
 Quant Title :
 QLast Update : Fri Oct 04 04:35:47 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.477	4711987	49.813 ug/ml
Target Compounds			
1) N-OCTANE	2.279	5860849	56.154 ug/ml
2) N-DECANE	4.343	6098238	55.970 ug/ml
3) N-DODECANE	6.338	6210269	55.372 ug/ml
4) N-TETRADECANE	8.109	6125767	54.063 ug/ml
5) N-HEXADECANE	9.690	6053840	52.996 ug/ml
6) N-OCTADECANE	11.114	5883404	52.112 ug/ml
7) N-EICOSANE	12.408	5779941	51.319 ug/ml
8) N-DOCOSANE	13.591	5563365	50.482 ug/ml
10) N-TETRACOSANE	14.678	5358918	49.774 ug/ml
11) N-HEXACOSANE	15.685	5163642	48.653 ug/ml
12) N-OCTACOSANE	16.621	5047302	45.386 ug/ml
13) N-TRIACONTANE	17.497	4890233	44.523 ug/ml
14) N-DOTRIACONTANE	18.316	4594728	47.675 ug/ml
15) N-TETRATRIACONTANE	19.091	4495238	50.490 ug/ml
16) N-HEXATRIACONTANE	19.822	4231650	51.040 ug/ml
17) N-OCTATRIACONTANE	20.520	3960796	52.823 ug/ml
18) N-TETRACONTANE	21.321	3963487	51.666 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG100319\
Data File : FG006132.D
Signal(s) : FID1A.ch
Acq On : 03 Oct 2019 16:59
Operator : DD\AJ
Sample : FG100319ICV
Misc :
ALS Vial : 8 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Oct 04 04:37:00 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG100319.M
Quant Title :
QLast Update : Fri Oct 04 04:35:47 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

