

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG100418\
 Data File : FG002504.D
 Signal(s) : FID1A.ch
 Acq On : 04 Oct 2018 09:21
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Oct 04 13:04:46 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG100118.M
 Quant Title :
 QLast Update : Mon Oct 01 12:36:17 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.762	6055218	52.754 ug/ml
Target Compounds			
2) N-DECANE	4.462	7121828	54.660 ug/ml
3) N-DODECANE	6.492	7227254	54.764 ug/ml
4) N-TETRADECANE	8.293	7298782	54.741 ug/ml
5) N-HEXADECANE	9.898	7339400	54.712 ug/ml
6) N-OCTADECANE	11.345	7345950	54.546 ug/ml
7) N-EICOSANE	12.658	7221619	54.139 ug/ml
8) N-DOCOSANE	13.860	7071035	53.587 ug/ml
10) N-TETRACOSANE	14.967	6847318	52.671 ug/ml
11) N-HEXACOSANE	15.990	6665191	51.887 ug/ml
12) N-OCTACOSANE	16.941	6510221	51.349 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG100418\
Data File : FG002504.D
Signal(s) : FID1A.ch
Acq On : 04 Oct 2018 09:21
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Oct 04 13:04:46 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG100118.M
Quant Title :
QLast Update : Mon Oct 01 12:36:17 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

