

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG100421\
 Data File : FG008767.D
 Signal(s) : FID1A.ch
 Acq On : 04 Oct 2021 14:01
 Operator : DD\AJ
 Sample : M4005-18MSD 10X
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Oct 05 01:14:59 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG090721.M
 Quant Title :
 QLast Update : Wed Sep 08 09:33:21 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.739	149015	1.342 ug/mlm
Target Compounds			
1) N-OCTANE	2.343	90163	0.683 ug/ml
2) N-DECANE	4.445	145077	1.085 ug/ml
3) N-DODECANE	6.470	157494	1.163 ug/ml
4) N-TETRADECANE	8.271	123176	0.908 ug/ml
5) N-HEXADECANE	9.876	130603	0.975 ug/ml
6) N-OCTADECANE	11.322	153435	1.140 ug/ml
7) N-EICOSANE	12.635	181229	1.362 ug/mlm
8) N-DOCOSANE	13.837	205519	1.594 ug/mlm
10) N-TETRACOSANE	14.943	188387	1.490 ug/mlm
11) N-HEXACOSANE	15.967	185847	1.512 ug/mlm
12) N-OCTACOSANE	16.921	252885	2.106 ug/ml
13) N-TRIACONTANE	17.811	343916	2.983 ug/ml
14) N-DOTRIACONTANE	18.648	303779	2.844 ug/ml
15) N-TETRATRIACONTANE	19.431	382721	3.917 ug/ml
16) N-HEXATRIACONTANE	20.171	258314	2.793 ug/mlm
17) N-OCTATRIACONTANE	20.930	372698	4.231 ug/mlm
18) N-TETRACONTANE	21.877	260689	2.801 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG100421\
 Data File : FG008767.D
 Signal(s) : FID1A.ch
 Acq On : 04 Oct 2021 14:01
 Operator : DD\AJ
 Sample : M4005-18MSD 10X
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Oct 05 01:14:59 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG090721.M
 Quant Title :
 QLast Update : Wed Sep 08 09:33:21 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

