

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG100518\
 Data File : FG002528.D
 Signal(s) : FID1A.ch
 Acq On : 05 Oct 2018 20:01
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Oct 06 00:44:46 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG100118.M
 Quant Title :
 QLast Update : Mon Oct 01 12:36:17 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.759	5777111	50.331 ug/ml
Target Compounds			
2) N-DECANE	4.463	6918602	53.100 ug/ml
3) N-DODECANE	6.494	6958193	52.725 ug/ml
4) N-TETRADECANE	8.294	6979651	52.347 ug/ml
5) N-HEXADECANE	9.899	6995375	52.147 ug/ml
6) N-OCTADECANE	11.345	6991107	51.911 ug/ml
7) N-EICOSANE	12.658	6875392	51.544 ug/ml
8) N-DOCOSANE	13.860	6751422	51.165 ug/ml
10) N-TETRACOSANE	14.964	6573247	50.563 ug/ml
11) N-HEXACOSANE	15.988	6449814	50.210 ug/ml
12) N-OCTACOSANE	16.939	6382860	50.344 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG100518\
Data File : FG002528.D
Signal(s) : FID1A.ch
Acq On : 05 Oct 2018 20:01
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Oct 06 00:44:46 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG100118.M
Quant Title :
QLast Update : Mon Oct 01 12:36:17 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

