

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG100518\
 Data File : FG002530.D
 Signal(s) : FID1A.ch
 Acq On : 05 Oct 2018 21:01
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Oct 06 00:45:11 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG100118.M
 Quant Title :
 QLast Update : Mon Oct 01 12:36:17 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.760	5692481	49.594 ug/ml
Target Compounds			
1) N-OCTANE	2.341	6379714	50.106 ug/ml
2) N-DECANE	4.463	6648157	51.025 ug/ml
3) N-DODECANE	6.493	6695224	50.732 ug/ml
4) N-TETRADECANE	8.293	6725344	50.440 ug/ml
5) N-HEXADECANE	9.899	6755839	50.362 ug/ml
6) N-OCTADECANE	11.344	6777600	50.326 ug/ml
7) N-EICOSANE	12.657	6697865	50.213 ug/ml
8) N-DOCOSANE	13.857	6612870	50.115 ug/ml
10) N-TETRACOSANE	14.964	6488653	49.912 ug/ml
11) N-HEXACOSANE	15.987	6412368	49.919 ug/ml
12) N-OCTACOSANE	16.938	6389643	50.397 ug/ml
13) N-TRIACONTANE	17.828	6485677	51.544 ug/ml
14) N-DOTRIACONTANE	18.661	6463553	52.647 ug/ml
15) N-TETRATRIACONTANE	19.448	6394737	53.591 ug/ml
16) N-HEXATRIACONTANE	20.189	6094914	54.936 ug/ml
17) N-OCTATRIACONTANE	20.951	5529082	56.406 ug/ml
18) N-TETRACONTANE	21.905	5465538	58.435 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG100518\
Data File : FG002530.D
Signal(s) : FID1A.ch
Acq On : 05 Oct 2018 21:01
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Oct 06 00:45:11 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG100118.M
Quant Title :
QLast Update : Mon Oct 01 12:36:17 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

