

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG100523\
 Data File : FG012993.D
 Signal(s) : FID1A.ch
 Acq On : 05 Oct 2023 12:57
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: Oct 05 23:54:41 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG092123.M
 Quant Title :
 QLast Update : Fri Sep 22 02:14:31 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.077	5770171	50.916 ug/ml
Target Compounds			
2) N-DECANE	4.595	6140279	53.355 ug/ml
3) N-DODECANE	6.768	6232679	53.078 ug/ml
4) N-TETRADECANE	8.602	6226637	52.714 ug/ml
5) N-HEXADECANE	10.214	6331203	52.333 ug/ml
6) N-OCTADECANE	11.661	6343005	51.979 ug/ml
7) N-EICOSANE	12.975	6316033	51.587 ug/ml
8) N-DOCOSANE	14.176	6111789	51.277 ug/ml
10) N-TETRACOSANE	15.281	6102831	50.924 ug/ml
11) N-HEXACOSANE	16.304	5989413	50.585 ug/ml
12) N-OCTACOSANE	17.255	5911682	50.503 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG100523\
Data File : FG012993.D
Signal(s) : FID1A.ch
Acq On : 05 Oct 2023 12:57
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
50 PPM TRPH STD

Integration File: autoint1.e
Quant Time: Oct 05 23:54:41 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG092123.M
Quant Title :
QLast Update : Fri Sep 22 02:14:31 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

