

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG100623\
 Data File : FG013011.D
 Signal(s) : FID1A.ch
 Acq On : 06 Oct 2023 11:53
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Oct 07 01:49:24 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG092123.M
 Quant Title :
 QLast Update : Fri Sep 22 02:14:31 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.076	5586169	49.292 ug/ml
Target Compounds			
1) N-OCTANE	2.074	5483754	49.039 ug/ml
2) N-DECANE	4.594	5792244	50.331 ug/ml
3) N-DODECANE	6.768	5929928	50.500 ug/ml
4) N-TETRADECANE	8.601	5957867	50.439 ug/ml
5) N-HEXADECANE	10.213	6077488	50.236 ug/ml
6) N-OCTADECANE	11.660	6113539	50.099 ug/ml
7) N-EICOSANE	12.973	6094319	49.776 ug/ml
8) N-DOCOSANE	14.175	5906678	49.556 ug/ml
10) N-TETRACOSANE	15.281	5909190	49.308 ug/ml
11) N-HEXACOSANE	16.303	5794744	48.941 ug/ml
12) N-OCTACOSANE	17.253	5715360	48.826 ug/ml
13) N-TRIACONTANE	18.143	5645083	48.902 ug/ml
14) N-DOTRIACONTANE	18.976	5415723	48.695 ug/ml
15) N-TETRATRIACONTANE	19.759	4648799	49.225 ug/ml
16) N-HEXATRIACONTANE	20.498	3680840	49.813 ug/ml
17) N-OCTATRIACONTANE	21.252	3290901	49.672 ug/ml
18) N-TETRACONTANE	22.198	3203963	50.810 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG100623\
Data File : FG013011.D
Signal(s) : FID1A.ch
Acq On : 06 Oct 2023 11:53
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Oct 07 01:49:24 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG092123.M
Quant Title :
QLast Update : Fri Sep 22 02:14:31 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

