

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG100623\
 Data File : FG013014.D
 Signal(s) : FID1A.ch
 Acq On : 06 Oct 2023 14:28
 Operator : YP\AJ
 Sample : PB156174BS
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Oct 07 01:49:49 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG092123.M
 Quant Title :
 QLast Update : Fri Sep 22 02:14:31 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.075	1696609	14.971 ug/ml
Target Compounds			
1) N-OCTANE	2.074	1757777	15.719 ug/ml
2) N-DECANE	4.593	2135096	18.553 ug/ml
3) N-DODECANE	6.766	2210145	18.822 ug/ml
4) N-TETRADECANE	8.599	2249664	19.045 ug/ml
5) N-HEXADECANE	10.211	2298750	19.001 ug/ml
6) N-OCTADECANE	11.659	2320771	19.018 ug/ml
7) N-EICOSANE	12.972	2309732	18.865 ug/ml
8) N-DOCOSANE	14.173	2255500	18.923 ug/ml
10) N-TETRACOSANE	15.279	2244436	18.728 ug/ml
11) N-HEXACOSANE	16.302	2205664	18.629 ug/ml
12) N-OCTACOSANE	17.253	2160315	18.455 ug/ml
13) N-TRIACONTANE	18.141	2129957	18.451 ug/ml
14) N-DOTRIACONTANE	18.975	2065545	18.572 ug/ml
15) N-TETRATRIACONTANE	19.757	1963609	20.792 ug/ml
16) N-HEXATRIACONTANE	20.497	1801989	24.386 ug/ml
17) N-OCTATRIACONTANE	21.251	1691081	25.525 ug/ml
18) N-TETRACONTANE	22.197	1646430	26.110 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG100623\
Data File : FG013014.D
Signal(s) : FID1A.ch
Acq On : 06 Oct 2023 14:28
Operator : YP\AJ
Sample : PB156174BS
Misc :
ALS Vial : 12 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Oct 07 01:49:49 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG092123.M
Quant Title :
QLast Update : Fri Sep 22 02:14:31 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

