

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG100623AL\
 Data File : FG013030.D
 Signal(s) : FID1A.ch
 Acq On : 07 Oct 2023 02:22
 Operator : YP\AJ
 Sample : 50 PPM ALIPHATIC HC STD2
 Misc :
 ALS Vial : 72 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Oct 07 07:42:51 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\Aliphatic EPH 100623.M
 Quant Title : GC Extractables
 QLast Update : Sat Oct 07 07:42:39 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.996	6266790	48.415 ug/ml
Spiked Amount 50.000		Recovery =	96.83%
12) S 1-chlorooctadecane (S...	13.435	4707704	48.450 ug/ml
Spiked Amount 50.000		Recovery =	96.90%
Target Compounds			
1) T n-Nonane (C9)	3.348	5516850	48.045 ug/ml
2) T n-Decane (C10)	4.602	5636960	48.087 ug/ml
3) T A~Naphthalene (C11.7)	6.314	6012498	48.123 ug/ml
4) T n-Dodecane (C12)	6.776	5664676	48.274 ug/ml
5) T A~2-methylnaphthalene...	7.415	5817354	48.283 ug/ml
6) T n-Tetradecane (C14)	8.611	5568184	48.370 ug/ml
7) T n-Hexadecane (C16)	10.225	5697194	48.412 ug/ml
8) T n-Octadecane (C18)	11.673	5688899	48.404 ug/ml
10) T n-Eicosane (C20)	12.988	5417448	48.425 ug/ml
11) T n-Heneicosane (C21)	13.602	5302384	48.440 ug/ml
13) T n-Docosane (C22)	14.190	5255153	48.440 ug/ml
14) T n-Tetracosane (C24)	15.296	5178990	48.349 ug/ml
15) T n-Hexacosane (C26)	16.320	5068453	48.332 ug/ml
16) T n-Octacosane (C28)	17.272	4961779	48.286 ug/ml
17) T n-Tricontane (C30)	18.161	5047580	47.991 ug/ml
18) T n-Dotriacontane (C32)	18.996	5001857	47.812 ug/ml
19) T n-Tetratriacontane (C34)	19.780	4970069	47.975 ug/ml
20) T n-Hexatriacontane (C36)	20.522	5043781	47.540 ug/ml
21) T n-Octatriacontane (C38)	21.283	5249717	48.415 ug/ml
22) T n-Tetracontane (C40)	22.238	5328785	47.946 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG100623AL\
Data File : FG013030.D
Signal(s) : FID1A.ch
Acq On : 07 Oct 2023 02:22
Operator : YP\AJ
Sample : 50 PPM ALIPHATIC HC STD2
Misc :
ALS Vial : 72 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Oct 07 07:42:51 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\Aliphatic EPH 100623.M
Quant Title : GC Extractables
QLast Update : Sat Oct 07 07:42:39 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um

