

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG100623AL\
 Data File : FG013031.D
 Signal(s) : FID1A.ch
 Acq On : 07 Oct 2023 02:51
 Operator : YP\AJ
 Sample : 20 PPM ALIPHATIC HC STD3
 Misc :
 ALS Vial : 73 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Oct 07 07:39:14 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\Aliphatic EPH 100623.M
 Quant Title : GC Extractables
 QLast Update : Sat Oct 07 07:38:53 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.991	2775215	20.000 ug/ml
Spiked Amount 50.000		Recovery =	40.00%
12) S 1-chlorooctadecane (S...	13.432	2080904	20.000 ug/ml
Spiked Amount 50.000		Recovery =	40.00%
Target Compounds			
1) T n-Nonane (C9)	3.346	2482749	20.000 ug/ml
2) T n-Decane (C10)	4.600	2534216	20.000 ug/ml
3) T A~Naphthalene (C11.7)	6.312	2684479	20.000 ug/ml
4) T n-Dodecane (C12)	6.774	2526172	20.000 ug/ml
5) T A~2-methylnaphthalene...	7.412	2581140	20.000 ug/ml
6) T n-Tetradecane (C14)	8.608	2472399	20.000 ug/ml
7) T n-Hexadecane (C16)	10.222	2525801	20.000 ug/ml
8) T n-Octadecane (C18)	11.670	2524134	20.000 ug/ml
10) T n-Eicosane (C20)	12.984	2402752	20.000 ug/ml
11) T n-Heneicosane (C21)	13.599	2349524	20.000 ug/ml
13) T n-Docosane (C22)	14.187	2331041	20.000 ug/ml
14) T n-Tetracosane (C24)	15.293	2306899	20.000 ug/ml
15) T n-Hexacosane (C26)	16.317	2261612	20.000 ug/ml
16) T n-Octacosane (C28)	17.269	2223711	20.000 ug/ml
17) T n-Tricontane (C30)	18.159	2312005	20.000 ug/ml
18) T n-Dotriacontane (C32)	18.993	2324439	20.000 ug/ml
19) T n-Tetratriacontane (C34)	19.778	2278541	20.000 ug/ml
20) T n-Hexatriacontane (C36)	20.517	2386730	20.000 ug/ml
21) T n-Octatriacontane (C38)	21.277	2348688	20.000 ug/ml
22) T n-Tetracontane (C40)	22.229	2428112	20.000 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG100623AL\
Data File : FG013031.D
Signal(s) : FID1A.ch
Acq On : 07 Oct 2023 02:51
Operator : YP\AJ
Sample : 20 PPM ALIPHATIC HC STD3
Misc :
ALS Vial : 73 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Oct 07 07:39:14 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\Aliphatic EPH 100623.M
Quant Title : GC Extractables
QLast Update : Sat Oct 07 07:38:53 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um

