

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG100818\
 Data File : FG002549.D
 Signal(s) : FID1A.ch
 Acq On : 08 Oct 2018 17:19
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Oct 09 02:28:35 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG100118.M
 Quant Title :
 QLast Update : Mon Oct 01 12:36:17 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.756	5168086	45.026 ug/ml
Target Compounds			
1) N-OCTANE	2.341	5935821	46.620 ug/ml
2) N-DECANE	4.462	6040432	46.360 ug/ml
3) N-DODECANE	6.491	6056515	45.893 ug/ml
4) N-TETRADECANE	8.292	6055145	45.413 ug/ml
5) N-HEXADECANE	9.897	6052128	45.116 ug/ml
6) N-OCTADECANE	11.343	6054693	44.958 ug/ml
7) N-EICOSANE	12.656	5983112	44.855 ug/ml
8) N-DOCOSANE	13.856	5921208	44.873 ug/ml
10) N-TETRACOSANE	14.961	5839624	44.920 ug/ml
11) N-HEXACOSANE	15.984	5793285	45.099 ug/ml
12) N-OCTACOSANE	16.936	5766052	45.479 ug/ml
13) N-TRIACONTANE	17.824	5794777	46.053 ug/ml
14) N-DOTRIACONTANE	18.658	5748303	46.821 ug/ml
15) N-TETRATRIACONTANE	19.444	5800882	48.614 ug/ml
16) N-HEXATRIACONTANE	20.184	5662811	51.041 ug/ml
17) N-OCTATRIACONTANE	20.947	5170412	52.747 ug/ml
18) N-TETRACONTANE	21.899	5116274	54.701 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG100818\
Data File : FG002549.D
Signal(s) : FID1A.ch
Acq On : 08 Oct 2018 17:19
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Oct 09 02:28:35 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG100118.M
Quant Title :
QLast Update : Mon Oct 01 12:36:17 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

