

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG100819\
 Data File : FG006144.D
 Signal(s) : FID1A.ch
 Acq On : 08 Oct 2019 16:43
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Oct 09 01:01:05 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG100319.M
 Quant Title :
 QLast Update : Fri Oct 04 04:35:47 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.474	4287196	45.323 ug/ml
Target Compounds			
2) N-DECANE	4.353	6118564	56.157 ug/ml
3) N-DODECANE	6.344	6113664	54.511 ug/ml
4) N-TETRADECANE	8.112	5889733	51.980 ug/ml
5) N-HEXADECANE	9.690	5709460	49.981 ug/ml
6) N-OCTADECANE	11.112	5470155	48.451 ug/ml
7) N-EICOSANE	12.405	5320480	47.239 ug/ml
8) N-DOCOSANE	13.587	5098984	46.268 ug/ml
10) N-TETRACOSANE	14.675	4898141	45.494 ug/ml
11) N-HEXACOSANE	15.681	4703960	44.321 ug/ml
12) N-OCTACOSANE	16.615	4527708	40.714 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG100819\
Data File : FG006144.D
Signal(s) : FID1A.ch
Acq On : 08 Oct 2019 16:43
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Oct 09 01:01:05 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG100319.M
Quant Title :
QLast Update : Fri Oct 04 04:35:47 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

