

rteres

Instrument :

FID_G

LabSampleId :

50 PPM TRPH STD

Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 10/12/2022

Supervised By :Ankita Jodhani 10/13/2022

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG10112
Data File : FG010858.D
Signal(s) : FID1A.ch
Acq On : 12 Oct 2022 03:50
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG092822.M
Title :

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	4.265	4.230	4.315	BB	608832	6500781	93.50%	8.932%
2	6.448	6.411	6.491	BB	636054	6616972	95.17%	9.091%
3	8.280	8.243	8.326	BB	608541	6688353	96.20%	9.189%
4	9.888	9.846	9.935	BB	604276	6858494	98.65%	9.423%
5	11.332	11.293	11.376	BB	584838	6949048	99.95%	9.548%
6	12.644	12.601	12.691	BB	555666	6952691	100.00%	9.553%
7	13.844	13.810	13.891	BB	535375	6779543	97.51%	9.315%
8	14.744	14.701	14.773	BV	409438	5696269	81.93%	7.826%
9	14.948	14.908	14.998	BB	513630	6721243	96.67%	9.235%
10	15.972	15.915	16.015	BB	501586	6569018	94.48%	9.025%
11	16.923	16.860	16.971	BB	462815	6451037	92.78%	8.863%
Sum of corrected areas:						72783447		

FG092822.M Wed Oct 12 08:42:12 2022