

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG101124\
 Data File : FG014412.D
 Signal(s) : FID1A.ch
 Acq On : 11 Oct 2024 14:26
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: Oct 14 03:01:29 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG100224.M
 Quant Title :
 QLast Update : Thu Oct 03 05:57:06 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.911	5803888	46.871 ug/ml
Target Compounds			
2) N-DECANE	4.432	6165911	52.344 ug/ml
3) N-DODECANE	6.606	6326512	54.061 ug/ml
4) N-TETRADECANE	8.438	6280944	51.724 ug/ml
5) N-HEXADECANE	10.049	6294936	49.453 ug/ml
6) N-OCTADECANE	11.495	6598978	49.370 ug/ml
7) N-EICOSANE	12.808	6682268	48.074 ug/ml
8) N-DOCOSANE	14.009	6499904	47.388 ug/ml
10) N-TETRACOSANE	15.116	6481530	47.019 ug/ml
11) N-HEXACOSANE	16.138	6447808	46.641 ug/ml
12) N-OCTACOSANE	17.090	6325961	46.163 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG101124\
Data File : FG014412.D
Signal(s) : FID1A.ch
Acq On : 11 Oct 2024 14:26
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
50 PPM TRPH STD

Integration File: autoint1.e
Quant Time: Oct 14 03:01:29 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG100224.M
Quant Title :
QLast Update : Thu Oct 03 05:57:06 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

