

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG101421\
 Data File : FG008880.D
 Signal(s) : FID1A.ch
 Acq On : 14 Oct 2021 22:36
 Operator : AJ\MA
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Oct 15 00:35:29 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG100721.M
 Quant Title :
 QLast Update : Fri Oct 08 03:31:08 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.743	6041507	52.341 ug/ml
Target Compounds			
1) N-OCTANE	2.340	6926564	50.142 ug/ml
2) N-DECANE	4.442	7040965	50.469 ug/ml
3) N-DODECANE	6.469	7072020	50.154 ug/ml
4) N-TETRADECANE	8.271	7106275	50.409 ug/ml
5) N-HEXADECANE	9.878	7070236	50.488 ug/ml
6) N-OCTADECANE	11.325	7202792	50.967 ug/ml
7) N-EICOSANE	12.639	7184788	51.518 ug/ml
8) N-DOCOSANE	13.842	6995880	51.867 ug/ml
10) N-TETRACOSANE	14.948	6855811	52.212 ug/ml
11) N-HEXACOSANE	15.972	6581139	52.571 ug/ml
12) N-OCTACOSANE	16.923	6283473	52.730 ug/ml
13) N-TRIACONTANE	17.814	5915019	52.150 ug/ml
14) N-DOTRIACONTANE	18.647	5435851	51.246 ug/ml
15) N-TETRATRIACONTANE	19.433	5067977	50.885 ug/ml
16) N-HEXATRIACONTANE	20.173	4757742	51.515 ug/mlm
17) N-OCTATRIACONTANE	20.928	4519876	51.492 ug/mlm
18) N-TETRACONTANE	21.870	4566098	51.279 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG101421\
Data File : FG008880.D
Signal(s) : FID1A.ch
Acq On : 14 Oct 2021 22:36
Operator : AJ\MA
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Oct 15 00:35:29 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG100721.M
Quant Title :
QLast Update : Fri Oct 08 03:31:08 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

