

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG101521\
 Data File : FG008898.D
 Signal(s) : FID1A.ch
 Acq On : 15 Oct 2021 20:42
 Operator : AJ\MA
 Sample : M4089-06MSD
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 20-SB-2-20211005-18.5-19.0MSD

Manual Integrations
 APPROVED

mohammad
 10/18/2021 4:37:33 PM

Integration File: autoint1.e
 Quant Time: Oct 16 00:53:38 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG100721.M
 Quant Title :
 QLast Update : Fri Oct 08 03:31:08 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.738	2114821	18.322 ug/ml
Target Compounds			
1) N-OCTANE	2.336	2239739	16.214 ug/ml
2) N-DECANE	4.439	2781353	19.936 ug/ml
3) N-DODECANE	6.466	2974943	21.098 ug/ml
4) N-TETRADECANE	8.267	3059298	21.701 ug/ml
5) N-HEXADECANE	9.873	3120468	22.283 ug/ml
6) N-OCTADECANE	11.320	3054132	21.611 ug/ml
7) N-EICOSANE	12.635	3171336	22.740 ug/ml
8) N-DOCOSANE	13.837	3241694	24.034 ug/ml
10) N-TETRACOSANE	14.942	4619052	35.178 ug/ml
11) N-HEXACOSANE	15.968	2958833	23.636 ug/ml
12) N-OCTACOSANE	16.920	3327962	27.928 ug/ml
13) N-TRIACONTANE	17.810	2707532	23.871 ug/mlm
14) N-DOTRIACONTANE	18.645	2478965	23.370 ug/mlm
15) N-TETRATRIACONTANE	19.428	1944443	19.523 ug/mlm
16) N-HEXATRIACONTANE	20.172	1546245	16.742 ug/mlm
17) N-OCTATRIACONTANE	20.925	1296608	14.771 ug/mlm
18) N-TETRACONTANE	21.869	1634292	18.354 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG101521\
Data File : FG008898.D
Signal(s) : FID1A.ch
Acq On : 15 Oct 2021 20:42
Operator : AJ\MA
Sample : M4089-06MSD
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
20-SB-2-20211005-18.5-19.0MSD

Manual Integrations
APPROVED

mohammad
10/18/2021 4:37:33 PM

Integration File: autoint1.e
Quant Time: Oct 16 00:53:38 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG100721.M
Quant Title :
QLast Update : Fri Oct 08 03:31:08 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

