

Instrument :
FID_G
ClientSampleId :
PB139921BS

rteres

Area Percent Report

Manual Integrations
APPROVED

mohammad
10/18/2021 4:37:25 PM

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG101521\
Data File : FG008895.D
Signal(s) : FID1A.ch
Acq On : 15 Oct 2021 19:09
Sample : PB139921BS
Misc :
ALS Vial : 10 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG100721.M
Title :

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.338	2.310	2.397	BB	260609	1912375	83.09%	5.856%
2	4.439	4.411	4.508	BB	242496	2219446	96.43%	6.797%
3	6.466	6.438	6.542	BB	219227	2250688	97.79%	6.893%
4	8.267	8.225	8.356	BB	202754	2240702	97.36%	6.862%
5	9.873	9.812	9.953	BB	193110	2211410	96.09%	6.772%
6	11.320	11.265	11.396	BB	192377	2266331	98.47%	6.940%
7	12.632	12.583	12.702	BB	194588	2301502	100.00%	7.048%
8	13.835	13.800	13.903	BB	181280	2269205	98.60%	6.949%
9	14.736	14.696	14.788	BB	139080	1908237	82.91%	5.844%
10	14.941	14.902	15.001	BB	180830	2168605	94.23%	6.641%
11	15.964	15.926	16.027	BB	154201	1992035	86.55%	6.100%
12	16.917	16.866	16.975	BB	138788	1841551	80.02%	5.640%
13	17.807	17.753	17.863	BB	127607	1730917	75.21%	5.301%
14	18.641	18.586	18.704	BB	113237	1637750	71.16%	5.015%
15	19.426	19.379	19.483	BB	79244	1180020	51.27%	3.614%
16	20.168	19.999	20.195	BV	60374	911986	39.63%	2.793%
17	20.920	20.854	20.992	BB	39529	738464	32.09%	2.261%
18	21.864	21.794	22.000	BBA	25319	872866	37.93%	2.673%
Sum of corrected areas:							32654091	

FG100721.M Mon Oct 18 17:28:14 2021