

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG101821\
Data File : FG008931.D
Signal(s) : FID1A.ch
Acq On : 18 Oct 2021 21:19
Operator : AJ\MA
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :

Integration File: autoint1.e
Quant Time: Oct 19 00:26:40 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG100721.M
Quant Title :
QLast Update : Fri Oct 08 03:31:08 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

	Compound	R.T.	Response	Conc Units

System Monitoring Compounds				
9) S	TETRACOSANE-d50 (SURR...	14.739	1822369	15.788 ug/ml
Target Compounds				
14)	N-DOTRIACONTANE	18.646	49295	0.465 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG101821\
Data File : FG008931.D
Signal(s) : FID1A.ch
Acq On : 18 Oct 2021 21:19
Operator : AJ\MA
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :

Integration File: autoint1.e
Quant Time: Oct 19 00:26:40 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG100721.M
Quant Title :
QLast Update : Fri Oct 08 03:31:08 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um






#9 TETRACOSANE-d50 (SURROGATE)

R.T.: 14.739 min
Delta R.T.: 0.003 min
Response: 1822369
Conc: 15.79 ug/ml

Instrument :
FID_G
ClientSampleId :



#14 N-DOTRIACONTANE

R.T.: 18.646 min
Delta R.T.: 0.006 min
Response: 49295
Conc: 0.46 ug/ml