

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG101921\
 Data File : FG008945.D
 Signal(s) : FID1A.ch
 Acq On : 19 Oct 2021 12:36
 Operator : AJ\MA
 Sample : M4068-12
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 MDL-WATER-06-QT4-2021

Manual Integrations
 APPROVED

mohammad
 10/20/2021 1:33:55 PM

Integration File: autoint1.e
 Quant Time: Oct 20 00:07:08 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG100721.M
 Quant Title :
 QLast Update : Fri Oct 08 03:31:08 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.736	1720790	14.908 ug/ml
Target Compounds			
1) N-OCTANE	2.340	125478	0.908 ug/ml
2) N-DECANE	4.439	105001	0.753 ug/ml
3) N-DODECANE	6.466	55402	0.393 ug/ml
4) N-TETRADECANE	8.282	71958	0.510 ug/mlm
5) N-HEXADECANE	9.880	82941	0.592 ug/ml
6) N-OCTADECANE	11.321	97810	0.692 ug/ml
7) N-EICOSANE	12.634	104733	0.751 ug/ml
8) N-DOCOSANE	13.833	108151	0.802 ug/ml
10) N-TETRACOSANE	14.941	107568	0.819 ug/ml
11) N-HEXACOSANE	15.971	97504	0.779 ug/ml
12) N-OCTACOSANE	16.921	98741	0.829 ug/ml
13) N-TRIACONTANE	17.809	115546	1.019 ug/ml
14) N-DOTRIACONTANE	18.643	139565	1.316 ug/ml
15) N-TETRATRIACONTANE	19.428	140574	1.411 ug/ml
16) N-HEXATRIACONTANE	20.171	168497	1.824 ug/mlm
17) N-OCTATRIACONTANE	20.923	157887	1.799 ug/ml
18) N-TETRACONTANE	21.868	236355	2.654 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG101921\
Data File : FG008945.D
Signal(s) : FID1A.ch
Acq On : 19 Oct 2021 12:36
Operator : AJ\MA
Sample : M4068-12
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
MDL-WATER-06-QT4-2021

Manual Integrations
APPROVED

mohammad
10/20/2021 1:33:55 PM

Integration File: autoint1.e
Quant Time: Oct 20 00:07:08 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG100721.M
Quant Title :
QLast Update : Fri Oct 08 03:31:08 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

