

rteres

Area Percent Report

Instrument :
FID_G
LabSampleId :
50 PPM TRPH STD

Manual Integrations
APPROVED

mohammad
10/20/2021 1:34:00 PM

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG101921\
Data File : FG008952.D
Signal(s) : FID1A.ch
Acq On : 19 Oct 2021 17:44
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG100721.M
Title :

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.336	2.305	2.396	BB	979641	7119345	98.22%	6.550%
2	4.438	4.411	4.512	BB	788186	7248022	100.00%	6.668%
3	6.466	6.434	6.545	BB	707628	7168612	98.90%	6.595%
4	8.268	8.235	8.359	BB	646982	7016865	96.81%	6.456%
5	9.873	9.840	9.960	BB	613676	6784452	93.60%	6.242%
6	11.320	11.282	11.397	BB	586147	6705660	92.52%	6.169%
7	12.633	12.593	12.706	BB	549188	6487279	89.50%	5.968%
8	13.836	13.802	13.904	VB	485226	6174523	85.19%	5.681%
9	14.738	14.686	14.768	BV	390771	5219454	72.01%	4.802%
10	14.943	14.870	15.005	BB	460324	5968361	82.34%	5.491%
11	15.966	15.915	16.032	BB	423198	5713540	78.83%	5.257%
12	16.919	16.854	16.980	BB	426945	5594949	77.19%	5.147%
13	17.808	17.752	17.870	BB	402645	5562902	76.75%	5.118%
14	18.642	18.585	18.711	BB	393225	5474459	75.53%	5.037%
15	19.429	19.360	19.489	BB	358796	5363192	74.00%	4.934%
16	20.171	20.129	20.221	M	345561	5148473	71.03%	4.737%
17	20.923	20.845	20.994	BB	282057	4948795	68.28%	4.553%
18	21.862	21.776	22.000	BBA	201663	4995346	68.92%	4.596%
Sum of corrected areas:						108694230		

FG100721.M Wed Oct 20 18:21:07 2021