

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG102023\
 Data File : FG013086.D
 Signal(s) : FID1A.ch
 Acq On : 20 Oct 2023 14:36
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Oct 21 00:47:29 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG101623.M
 Quant Title :
 QLast Update : Mon Oct 16 12:31:23 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.092	5444915	51.977 ug/ml
Target Compounds			
1) N-OCTANE	2.076	5384045	47.345 ug/ml
2) N-DECANE	4.600	5861408	50.981 ug/ml
3) N-DODECANE	6.776	5973231	51.713 ug/ml
4) N-TETRADECANE	8.611	5970589	52.362 ug/ml
5) N-HEXADECANE	10.225	6078403	52.347 ug/ml
6) N-OCTADECANE	11.673	6064048	52.224 ug/ml
7) N-EICOSANE	12.989	5999973	52.124 ug/ml
8) N-DOCOSANE	14.190	5776706	52.206 ug/ml
10) N-TETRACOSANE	15.297	5762445	52.447 ug/ml
11) N-HEXACOSANE	16.321	5650747	52.509 ug/ml
12) N-OCTACOSANE	17.273	5606916	52.585 ug/ml
13) N-TRIACONTANE	18.163	5680980	53.043 ug/ml
14) N-DOTRIACONTANE	18.997	5480794	52.362 ug/ml
15) N-TETRATRIACONTANE	19.782	4881790	53.634 ug/ml
16) N-HEXATRIACONTANE	20.521	3972440	54.010 ug/ml
17) N-OCTATRIACONTANE	21.280	3702211	52.308 ug/mlm
18) N-TETRACONTANE	22.236	3718291	51.296 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG102023\
Data File : FG013086.D
Signal(s) : FID1A.ch
Acq On : 20 Oct 2023 14:36
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Oct 21 00:47:29 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG101623.M
Quant Title :
QLast Update : Mon Oct 16 12:31:23 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

