

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG102122\
 Data File : FG010960.D
 Signal(s) : FID1A.ch
 Acq On : 21 Oct 2022 21:09
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Oct 22 16:00:19 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG092822.M
 Quant Title :
 QLast Update : Thu Sep 29 02:21:49 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
9) S	TETRACOSANE-d50 (SURR...	14.742	4922509	40.575 ug/ml
Target Compounds				
2)	N-DECANE	4.261	5804507	44.603 ug/ml
3)	N-DODECANE	6.444	5864495	44.220 ug/ml
4)	N-TETRADECANE	8.276	5854291	43.659 ug/ml
5)	N-HEXADECANE	9.885	5933967	42.796 ug/ml
6)	N-OCTADECANE	11.329	5960046	42.289 ug/ml
7)	N-EICOSANE	12.640	5886160	41.625 ug/ml
8)	N-DOCOSANE	13.840	5675193	40.996 ug/ml
10)	N-TETRACOSANE	14.947	5560370	40.359 ug/ml
11)	N-HEXACOSANE	15.969	5411835	40.114 ug/ml
12)	N-OCTACOSANE	16.921	5288214	40.327 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG102122\
Data File : FG010960.D
Signal(s) : FID1A.ch
Acq On : 21 Oct 2022 21:09
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Oct 22 16:00:19 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG092822.M
Quant Title :
QLast Update : Thu Sep 29 02:21:49 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

