

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG102422\
 Data File : FG010973.D
 Signal(s) : FID1A.ch
 Acq On : 24 Oct 2022 11:17
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: Oct 25 06:05:30 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG092822.M
 Quant Title :
 QLast Update : Thu Sep 29 02:21:49 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.739	5403217	44.537 ug/ml
Target Compounds			
2) N-DECANE	4.259	6356864	48.847 ug/ml
3) N-DODECANE	6.443	6415413	48.374 ug/ml
4) N-TETRADECANE	8.275	6391100	47.662 ug/ml
5) N-HEXADECANE	9.883	6472009	46.676 ug/ml
6) N-OCTADECANE	11.327	6505700	46.160 ug/ml
7) N-EICOSANE	12.640	6421275	45.409 ug/ml
8) N-DOCOSANE	13.838	6186858	44.693 ug/ml
10) N-TETRACOSANE	14.945	6099545	44.273 ug/ml
11) N-HEXACOSANE	15.967	5892664	43.678 ug/ml
12) N-OCTACOSANE	16.920	5769621	43.998 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG102422\
Data File : FG010973.D
Signal(s) : FID1A.ch
Acq On : 24 Oct 2022 11:17
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
50 PPM TRPH STD

Integration File: autoint1.e
Quant Time: Oct 25 06:05:30 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG092822.M
Quant Title :
QLast Update : Thu Sep 29 02:21:49 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

