

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG102720\
 Data File : FG007250.D
 Signal(s) : FID1A.ch
 Acq On : 27 Oct 2020 11:20
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: Oct 27 15:43:10 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG100220.M
 Quant Title :
 QLast Update : Mon Oct 05 13:04:52 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.754	6224529	51.120 ug/ml
Target Compounds			
2) N-DECANE	4.449	8070440	59.483 ug/ml
3) N-DODECANE	6.480	8058259	58.799 ug/ml
4) N-TETRADECANE	8.282	7926481	58.078 ug/ml
5) N-HEXADECANE	9.888	7959360	56.516 ug/ml
6) N-OCTADECANE	11.336	7835057	55.019 ug/ml
7) N-EICOSANE	12.650	7574685	53.547 ug/ml
8) N-DOCOSANE	13.853	7287059	52.204 ug/ml
10) N-TETRACOSANE	14.960	7102011	50.905 ug/ml
11) N-HEXACOSANE	15.984	6826801	49.832 ug/ml
12) N-OCTACOSANE	16.936	6645908	49.276 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG102720\
Data File : FG007250.D
Signal(s) : FID1A.ch
Acq On : 27 Oct 2020 11:20
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
50 PPM TRPH STD

Integration File: autoint1.e
Quant Time: Oct 27 15:43:10 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG100220.M
Quant Title :
QLast Update : Mon Oct 05 13:04:52 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

