

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG102922\
 Data File : FG011063.D
 Signal(s) : FID1A.ch
 Acq On : 28 Oct 2022 11:30
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: Oct 29 07:14:52 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG102722.M
 Quant Title :
 QLast Update : Thu Oct 27 00:27:30 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.728	5356396	43.540 ug/ml
Target Compounds			
2) N-DECANE	4.251	6400660	44.140 ug/ml
3) N-DODECANE	6.434	6503241	44.539 ug/ml
4) N-TETRADECANE	8.265	6478455	45.184 ug/ml
5) N-HEXADECANE	9.873	6540949	44.537 ug/ml
6) N-OCTADECANE	11.316	6532232	43.910 ug/ml
7) N-EICOSANE	12.627	6385993	43.653 ug/ml
8) N-DOCOSANE	13.826	6103471	43.363 ug/ml
10) N-TETRACOSANE	14.931	6022409	43.571 ug/ml
11) N-HEXACOSANE	15.954	5861908	43.180 ug/ml
12) N-OCTACOSANE	16.905	5827974	43.112 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG102922\
Data File : FG011063.D
Signal(s) : FID1A.ch
Acq On : 28 Oct 2022 11:30
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
50 PPM TRPH STD

Integration File: autoint1.e
Quant Time: Oct 29 07:14:52 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG102722.M
Quant Title :
QLast Update : Thu Oct 27 00:27:30 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

