

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG103019\
 Data File : FG006214.D
 Signal(s) : FID1A.ch
 Acq On : 31 Oct 2019 02:22
 Operator : DD\AJ
 Sample : 5 TRPH STD
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 5 TRPH STD

Manual Integrations
 APPROVED

Ankita
 10/31/2019 9:51:17 AM

Integration File: autoint1.e
 Quant Time: Oct 31 03:24:38 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG103019.M
 Quant Title :
 QLast Update : Thu Oct 31 03:14:45 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.466	398763	4.888 ug/ml
Target Compounds			
1) N-OCTANE	2.312	465225	4.680 ug/ml
2) N-DECANE	4.367	479370	4.682 ug/ml
3) N-DODECANE	6.353	484225	4.689 ug/ml
4) N-TETRADECANE	8.116	487118	4.743 ug/ml
5) N-HEXADECANE	9.690	488051	4.758 ug/ml
6) N-OCTADECANE	11.109	484009	4.832 ug/ml
7) N-EICOSANE	12.399	482891	4.891 ug/ml
8) N-DOCOSANE	13.580	468683	4.903 ug/ml
10) N-TETRACOSANE	14.667	459297	4.936 ug/ml
11) N-HEXACOSANE	15.672	449159	4.950 ug/ml
12) N-OCTACOSANE	16.608	459408	5.177 ug/ml
13) N-TRIACONTANE	17.482	482939	5.579 ug/ml
14) N-DOTRIACONTANE	18.301	477632	5.835 ug/ml
15) N-TETRATRIACONTANE	19.073	397517	5.101 ug/ml
16) N-HEXATRIACONTANE	19.802	437717	6.173 ug/ml
17) N-OCTATRIACONTANE	20.496	312661	4.820 ug/mlm
18) N-TETRACONTANE	21.285	363244	5.652 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG103019\
Data File : FG006214.D
Signal(s) : FID1A.ch
Acq On : 31 Oct 2019 02:22
Operator : DD\AJ
Sample : 5 TRPH STD
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
5 TRPH STD

Manual Integrations
APPROVED

Ankita
10/31/2019 9:51:17 AM

Integration File: autoint1.e
Quant Time: Oct 31 03:24:38 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG103019.M
Quant Title :
QLast Update : Thu Oct 31 03:14:45 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

