

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG103122\
 Data File : FG011086.D
 Signal(s) : FID1A.ch
 Acq On : 31 Oct 2022 23:51
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: Nov 01 07:40:57 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG102722.M
 Quant Title :
 QLast Update : Thu Oct 27 00:27:30 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.720	5472782	44.486 ug/ml
Target Compounds			
2) N-DECANE	4.246	6599103	45.509 ug/ml
3) N-DODECANE	6.429	6667359	45.663 ug/ml
4) N-TETRADECANE	8.261	6624170	46.200 ug/ml
5) N-HEXADECANE	9.868	6671719	45.427 ug/ml
6) N-OCTADECANE	11.311	6634613	44.598 ug/ml
7) N-EICOSANE	12.621	6479101	44.290 ug/ml
8) N-DOCOSANE	13.820	6197700	44.032 ug/ml
10) N-TETRACOSANE	14.924	6137850	44.406 ug/ml
11) N-HEXACOSANE	15.947	5974455	44.009 ug/ml
12) N-OCTACOSANE	16.898	5976115	44.208 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG103122\
Data File : FG011086.D
Signal(s) : FID1A.ch
Acq On : 31 Oct 2022 23:51
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
50 PPM TRPH STD

Integration File: autoint1.e
Quant Time: Nov 01 07:40:57 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG102722.M
Quant Title :
QLast Update : Thu Oct 27 00:27:30 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

