

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG110221\
 Data File : FG008983.D
 Signal(s) : FID1A.ch
 Acq On : 02 Nov 2021 11:50
 Operator : AJ\MA
 Sample : PB140348BS
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 PB140348BS

Integration File: autoint1.e
 Quant Time: Nov 03 01:06:49 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG110121.M
 Quant Title :
 QLast Update : Tue Nov 02 01:56:13 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.729	2248138	19.478 ug/ml
Target Compounds			
2) N-DECANE	4.435	2645265	19.564 ug/ml
3) N-DODECANE	6.460	2706118	19.930 ug/ml
4) N-TETRADECANE	8.262	2727243	20.534 ug/ml
5) N-HEXADECANE	9.867	2719404	20.513 ug/ml
6) N-OCTADECANE	11.313	2783696	20.302 ug/ml
7) N-EICOSANE	12.626	2804651	20.505 ug/ml
8) N-DOCOSANE	13.829	2760326	20.625 ug/ml
10) N-TETRACOSANE	14.935	2663233	20.162 ug/ml
11) N-HEXACOSANE	15.958	2481539	19.490 ug/ml
12) N-OCTACOSANE	16.910	2316044	18.839 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG110221\
Data File : FG008983.D
Signal(s) : FID1A.ch
Acq On : 02 Nov 2021 11:50
Operator : AJ\MA
Sample : PB140348BS
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
PB140348BS

Integration File: autoint1.e
Quant Time: Nov 03 01:06:49 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG110121.M
Quant Title :
QLast Update : Tue Nov 02 01:56:13 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

