

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG110221\
 Data File : FG008988.D
 Signal(s) : FID1A.ch
 Acq On : 02 Nov 2021 15:07
 Operator : AJ\MA
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: Nov 03 01:08:13 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG110121.M
 Quant Title :
 QLast Update : Tue Nov 02 01:56:13 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.732	6317064	54.732 ug/ml
Target Compounds			
2) N-DECANE	4.435	7562698	55.932 ug/ml
3) N-DODECANE	6.462	7620809	56.125 ug/ml
4) N-TETRADECANE	8.264	7644563	57.557 ug/ml
5) N-HEXADECANE	9.870	7548233	56.939 ug/ml
6) N-OCTADECANE	11.315	7624118	55.605 ug/ml
7) N-EICOSANE	12.629	7534761	55.087 ug/ml
8) N-DOCOSANE	13.831	7331613	54.783 ug/ml
10) N-TETRACOSANE	14.938	7231153	54.744 ug/ml
11) N-HEXACOSANE	15.961	6977364	54.801 ug/ml
12) N-OCTACOSANE	16.914	6712582	54.602 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG110221\
Data File : FG008988.D
Signal(s) : FID1A.ch
Acq On : 02 Nov 2021 15:07
Operator : AJ\MA
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
50 PPM TRPH STD

Integration File: autoint1.e
Quant Time: Nov 03 01:08:13 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG110121.M
Quant Title :
QLast Update : Tue Nov 02 01:56:13 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

