

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG110920\
 Data File : FG007325.D
 Signal(s) : FID1A.ch
 Acq On : 09 Nov 2020 15:46
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Nov 10 00:34:14 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG110220.M
 Quant Title :
 QLast Update : Tue Nov 03 02:04:00 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.748	5148267	41.847 ug/ml
Target Compounds			
2) N-DECANE	4.446	6493165	47.125 ug/ml
3) N-DODECANE	6.476	6516887	46.528 ug/ml
4) N-TETRADECANE	8.278	6453941	45.770 ug/ml
5) N-HEXADECANE	9.885	6412284	44.824 ug/ml
6) N-OCTADECANE	11.332	6353842	43.887 ug/ml
7) N-EICOSANE	12.646	6168230	43.011 ug/ml
8) N-DOCOSANE	13.847	5935623	42.276 ug/ml
10) N-TETRACOSANE	14.955	5819752	41.683 ug/ml
11) N-HEXACOSANE	15.978	5638270	41.419 ug/ml
12) N-OCTACOSANE	16.930	5516220	41.540 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG110920\
Data File : FG007325.D
Signal(s) : FID1A.ch
Acq On : 09 Nov 2020 15:46
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Nov 10 00:34:14 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG110220.M
Quant Title :
QLast Update : Tue Nov 03 02:04:00 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

