

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG111218\
 Data File : FG002905.D
 Signal(s) : FID1A.ch
 Acq On : 13 Nov 2018 01:46
 Operator : AJ\SJ
 Sample : J5887-14 10X
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Nov 13 03:18:57 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG102618.M
 Quant Title :
 QLast Update : Sat Oct 27 02:13:37 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.754	141030	1.283 ug/ml
Target Compounds			
7) N-EICOSANE	12.721f	197416	1.523 ug/ml
10) N-TETRACOSANE	14.947	110874	0.888 ug/ml
12) N-OCTACOSANE	16.935	14094	0.116 ug/ml
13) N-TRIACONTANE	17.824	179638	1.497 ug/ml
14) N-DOTRIACONTANE	18.662	183557	1.592 ug/ml
15) N-TETRATRIACONTANE	19.446	140567	1.264 ug/ml
16) N-HEXATRIACONTANE	20.195	479560	4.554 ug/ml
17) N-OCTATRIACONTANE	20.952	143792	1.511 ug/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG111218\
Data File : FG002905.D
Signal(s) : FID1A.ch
Acq On : 13 Nov 2018 01:46
Operator : AJ\SJ
Sample : J5887-14 10X
Misc :
ALS Vial : 26 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Nov 13 03:18:57 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG102618.M
Quant Title :
QLast Update : Sat Oct 27 02:13:37 2018
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um






#7 N-EICOSANE

R.T.: 12.721 min
Delta R.T.: 0.065 min
Response: 197416
Conc: 1.52 ug/ml




#9 TETRACOSANE-d50 (SURROGATE)

R.T.: 14.754 min
Delta R.T.: -0.009 min
Response: 141030
Conc: 1.28 ug/ml




#10 N-TETRACOSANE

R.T.: 14.947 min
Delta R.T.: -0.019 min
Response: 110874
Conc: 0.89 ug/ml




#12 N-OCTACOSANE

R.T.: 16.935 min
Delta R.T.: -0.009 min
Response: 14094
Conc: 0.12 ug/ml




#13 N-TRIACONTANE

R.T.: 17.824 min
Delta R.T.: -0.011 min
Response: 179638
Conc: 1.50 ug/ml




#14 N-DOTRIACONTANE

R.T.: 18.662 min
Delta R.T.: -0.011 min
Response: 183557
Conc: 1.59 ug/ml




#15 N-TETRATRIACONTANE

R.T.: 19.446 min
Delta R.T.: -0.012 min
Response: 140567
Conc: 1.26 ug/ml



#16 N-HEXATRIACONTANE

R.T.: 20.195 min
Delta R.T.: -0.006 min
Response: 479560
Conc: 4.55 ug/ml



#17 N-OCTATRIACONTANE

R.T.: 20.952 min
Delta R.T.: -0.017 min
Response: 143792
Conc: 1.51 ug/ml