

rteres

Instrument :
FID_G
LabSampleId :
50 PPM TRPH STD
Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Ankita Jodhani 11/15/2021
Supervised By :mohammad ahmed 11/15/2021

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG11122
Data File : FG009059.D
Signal(s) : FID1A.ch
Acq On : 13 Nov 2021 04:30
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG110121.M
Title :

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.332	2.300	2.394	BB	963678	6981568	98.79%	6.310%
2	4.436	4.408	4.517	BB	778538	7066775	100.00%	6.387%
3	6.464	6.431	6.549	BB	698418	7017863	99.31%	6.343%
4	8.267	8.234	8.360	BB	636466	6895424	97.58%	6.232%
5	9.875	9.838	9.973	BB	615309	6735369	95.31%	6.088%
6	11.322	11.278	11.406	BB	583441	6764988	95.73%	6.114%
7	12.638	12.594	12.708	BB	555046	6642566	94.00%	6.004%
8	13.841	13.769	13.911	BB	513620	6429581	90.98%	5.811%
9	14.745	14.693	14.778	BV	392116	5498566	77.81%	4.970%
10	14.949	14.901	15.011	BB	490662	6282015	88.90%	5.678%
11	15.974	15.925	16.040	BB	468578	6067108	85.85%	5.484%
12	16.928	16.866	16.999	BB	445482	5912055	83.66%	5.343%
13	17.819	17.763	17.878	BB	411565	5788984	81.92%	5.232%
14	18.655	18.596	18.722	BB	401277	5623982	79.58%	5.083%
15	19.442	19.262	19.515	BB	383142	5537194	78.36%	5.005%
16	20.184	19.889	20.252	BB	362680	5239742	74.15%	4.736%
17	20.938	20.869	21.022	BB	290830	5079168	71.87%	4.591%
18	21.884	21.796	22.000	BBA	209849	5078838	71.87%	4.590%
Sum of corrected areas:						110641785		

FG110121.M Mon Nov 15 09:14:41 2021