

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG111318\
 Data File : FG002920.D
 Signal(s) : FID1A.ch
 Acq On : 13 Nov 2018 17:28
 Operator : AJ\SJ
 Sample : J5909-02
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Nov 14 03:54:36 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG102618.M
 Quant Title :
 QLast Update : Sat Oct 27 02:13:37 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.752	1851761	16.849 ug/ml
Target Compounds			
4) N-TETRADECANE	8.348f	384839	2.910 ug/ml
7) N-EICOSANE	12.599f	61699	0.476 ug/ml
8) N-DOCOSANE	13.849	114580	0.900 ug/ml
10) N-TETRACOSANE	14.941	938627	7.513 ug/ml
11) N-HEXACOSANE	15.978	147306	1.195 ug/ml
12) N-OCTACOSANE	16.943	702512	5.776 ug/ml
13) N-TRIACONTANE	17.818	514971	4.292 ug/ml
14) N-DOTRIACONTANE	18.652	255060	2.213 ug/ml
15) N-TETRATRIACONTANE	19.511f	158115	1.422 ug/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG111318\
Data File : FG002920.D
Signal(s) : FID1A.ch
Acq On : 13 Nov 2018 17:28
Operator : AJ\SJ
Sample : J5909-02
Misc :
ALS Vial : 11 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Nov 14 03:54:36 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG102618.M
Quant Title :
QLast Update : Sat Oct 27 02:13:37 2018
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um






#4 N-TETRADECANE

R.T.: 8.348 min
Delta R.T.: 0.058 min
Response: 384839
Conc: 2.91 ug/ml




#7 N-EICOSANE

R.T.: 12.599 min
Delta R.T.: -0.058 min
Response: 61699
Conc: 0.48 ug/ml




#8 N-DOCOSANE

R.T.: 13.849 min
Delta R.T.: -0.010 min
Response: 114580
Conc: 0.90 ug/ml




#9 TETRACOSANE-d50 (SURROGATE)

R.T.: 14.752 min
Delta R.T.: -0.011 min
Response: 1851761
Conc: 16.85 ug/ml




#10 N-TETRACOSANE

R.T.: 14.941 min
Delta R.T.: -0.025 min
Response: 938627
Conc: 7.51 ug/ml




#11 N-HEXACOSANE

R.T.: 15.978 min
Delta R.T.: -0.013 min
Response: 147306
Conc: 1.20 ug/ml




#12 N-OCTACOSANE

R.T.: 16.943 min
Delta R.T.: -0.001 min
Response: 702512
Conc: 5.78 ug/ml




#13 N-TRIACONTANE

R.T.: 17.818 min
Delta R.T.: -0.018 min
Response: 514971
Conc: 4.29 ug/ml



#14 N-DOTRIACONTANE

R.T.: 18.652 min
Delta R.T.: -0.020 min
Response: 255060
Conc: 2.21 ug/ml



#15 N-TETRATRIACONTANE

R.T.: 19.511 min
Delta R.T.: 0.053 min
Response: 158115
Conc: 1.42 ug/ml