

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG111418\
 Data File : FG002948.D
 Signal(s) : FID1A.ch
 Acq On : 14 Nov 2018 11:02
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: Sample.e
 Quant Time: Nov 14 15:10:57 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG102618.M
 Quant Title :
 QLast Update : Sat Oct 27 02:13:37 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.750	5318703	48.396 ug/ml
Target Compounds			
2) N-DECANE	4.460	6653288	52.165 ug/ml
3) N-DODECANE	6.490	6635845	50.912 ug/ml
4) N-TETRADECANE	8.288	6609462	49.987 ug/ml
5) N-HEXADECANE	9.892	6526094	49.009 ug/ml
6) N-OCTADECANE	11.337	6426361	48.579 ug/ml
7) N-EICOSANE	12.649	6252397	48.231 ug/ml
8) N-DOCOSANE	13.849	6114082	48.029 ug/ml
10) N-TETRACOSANE	14.955	6010610	48.113 ug/ml
11) N-HEXACOSANE	15.978	5734856	46.532 ug/mlm
12) N-OCTACOSANE	16.928	5721631	47.042 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG111418\
Data File : FG002948.D
Signal(s) : FID1A.ch
Acq On : 14 Nov 2018 11:02
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: Sample.e
Quant Time: Nov 14 15:10:57 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG102618.M
Quant Title :
QLast Update : Sat Oct 27 02:13:37 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

