

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG111418\
 Data File : FG002962.D
 Signal(s) : FID1A.ch
 Acq On : 14 Nov 2018 18:31
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: Sample.e
 Quant Time: Nov 15 01:13:29 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG102618.M
 Quant Title :
 QLast Update : Sat Oct 27 02:13:37 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.749	5200955	47.324 ug/ml
Target Compounds			
2) N-DECANE	4.460	7067689	55.415 ug/ml
3) N-DODECANE	6.489	6982555	53.572 ug/ml
4) N-TETRADECANE	8.288	6890291	52.110 ug/ml
5) N-HEXADECANE	9.892	6766786	50.816 ug/ml
6) N-OCTADECANE	11.336	6615935	50.012 ug/ml
7) N-EICOSANE	12.648	6378210	49.202 ug/ml
8) N-DOCOSANE	13.848	6160969	48.397 ug/ml
10) N-TETRACOSANE	14.954	5978492	47.856 ug/ml
11) N-HEXACOSANE	15.976	5878644	47.699 ug/ml
12) N-OCTACOSANE	16.927	5897114	48.485 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG111418\
Data File : FG002962.D
Signal(s) : FID1A.ch
Acq On : 14 Nov 2018 18:31
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: Sample.e
Quant Time: Nov 15 01:13:29 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG102618.M
Quant Title :
QLast Update : Sat Oct 27 02:13:37 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

